



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6
RECEIVED 7/11/14-CD

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED

By Carol Day at 3:36 pm, Aug 12, 2014

DATAMASTER SN 204177	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Troop D Satellite, Carthage, Missouri		TIME OF INSPECTION 9:46 pm
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.		
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 07-70-14 9:46pm
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER	49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>GUTH LABORATORIES</u> LOT # <u>13290</u> EXP. DATE <u>10/29/2015</u>		
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34</u> °C SIMULATOR SN <u>G6754</u> EXP. DATE <u>01/28/2015</u>		
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		
TEST 1 • .101	TEST 2 • .103	TEST 3 • .103
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)		
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)		
REFUSALS 0	(0-.04) 5	(.05-.09) 0
(.10-.14) 1	(.15-.19) 0	OVER .19 0
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY). This unit is operating within the rules and regulations of the Missouri Department of Health and Senior Services.		
INSPECTING OFFICER		
SIGNATURE 	PRINT FULL NAME Jeffery L. Prewitt	
TYPE II PERMIT NUMBER/EXPIRATION DATE 240067 03/07/2016	TELEPHONE NUMBER (417) 895-6868	
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901		



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
07/07/14

ARREST TIME: 21:00
SUBJECT NAME:

TEST

DOB: 09/09/77 SEX: M

STATE/D.L.: MO/NONE

ARRESTING OFFICER:

PERMIT

OFFICER I.D.: 981

TESTING OFFICER:

SEWE

OFFICER I.D.: 981

PERMIT NUMBER: 240057

EXPIRATION DATE: 03/07/16

MISCELLANEOUS DATA:

TEST

--- PERMIT ANALYSIS ---

BLANK TEST .000 21:45
INTERNAL STANDARD VERIFIED
RADIO INTERFERENCE 21:50

Operator Signature

[Signature]

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
07/07/14

21:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

RETEERS
SERIAL CHAMBER: 49C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

INTERNAL STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

*****012345678910111213141516171819
ABCDEFGHIJKLMN O PQRSTUVWXYZ[\] ^ _ `abcde fghijklmno
pqrstuvwxyz{|}~

Operator Signature

[Signature]

Face This Side Down - This Edge In First

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BRO DETRANMASTER SERIAL NUMBER 284177
07/07/14

TESTING OFFICER:
PREWITT/J/L
OFFICER I.D.# 981
PERMIT NUMBER: 240067
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
TEST

--- SUPERVISOR NOTE ---

BLANK TEST	.000	02:16
INTERNAL STANDARD	VERIFIED	02:16
EXTERNAL STANDARD	.101	02:17
BLANK TEST	.000	02:17
EXTERNAL STANDARD	.103	02:18
BLANK TEST	.000	02:18
EXTERNAL STANDARD	.103	02:19
BLANK TEST	.000	02:20

N = 3
STDEV. = .1
AVG. = .1023

Operator Signature Q T Hurd



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JEFFERY L PREWITT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240067

EXPIRES 3/7/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PREWITT, JEFFERY
Permit No 240067
Date Issued 3/7/2014 Date Expires 3/7/2016